

MI 3123 Earth / Clamp

SMART^{TEC}[®]



Professional
solution for
smart
operator



Main Features:

- Supports all types of earthing resistance measurements:
 - **Standard earthing resistance** measurement with spikes
 - **Selective earthing resistance** test with one current clamp
 - Selective (contact less) **earthing resistance with two current clamps**
 - **Specific earth resistance** measurement
- Leakage current tests with current clamps
- Load current tests with current clamps
- **High accuracy, high immunity against external noise**
- **Preprogrammed limits**
- PASS/FAIL evaluation of measurement result
- **Bright GREEN/RED indicators** provides visual evaluation of results
- Suited for testing on **CAT IV** installations
- **Magnetic holder** enables hands-free operation
- **Built-in battery charger**
- Two level memory structure for saving test results and parameters
- Basic downloading with EuroLink PRO PC SW package
- Compatible with EuroLink PRO Plus PC SW package
- **USB and RS232** ports
- Robust and ergonomic design



Two clamps method for testing ground resistance of object on TN system



Large LCD screen with backlight and PASS/FAIL indicators

Suitable for:

- Testing earthing resistance
- Testing main earthing of objects
- Testing resistance of individual earthing rods in large systems
- Testing lightning systems
- Determination of soil resistivity
- Designing of earthing systems
- Testing on CAT IV installations (outside installations, main distribution boards, industrial plants)

Standards:

Instrument:
IEC/EN 61557 Parts 1,5,10
IEC/EN 61010-1
IEC/EN 61326

Applications:
IEC/EN 60364
VDE 100
BS 7671 17th edition
CEI 64.8

Technical specifications

Earth resistance

Earth resistance 4-wire method and Earth resistance one clamp method
Measuring range according to EN61557 is $0.67 \Omega \div 9999 \Omega$

Measuring range R (Ω)	Resolution (Ω)	Accuracy
$0.00 \div 19.99$	0.01	$\pm(3 \% \text{ of reading} + 3 \text{ digits})$
$20.0 \div 199.9$	0.1	$\pm(3 \% \text{ of reading} + 3 \text{ digits})$
$200 \div 1999$	1	$\pm 5 \% \text{ of reading}$
$2000 \div 9999$	1	$\pm 10 \% \text{ of reading}$

Automatic test of probe resistance	yes
Automatic test of voltage noise	yes
Open-terminal test voltage	40 VAC
Test voltage frequency	125 Hz
Short-circuit test current	< 20 mA
Low clamp current indication	yes
Noise current indication	yes

Earth resistance two clamps method

Measuring range (Ω)	Resolution (Ω)	Accuracy
$0.00 \div 19.9$	0.01	$\pm(10 \% \text{ of reading} + 10 \text{ digits})$
$20.0 \div 30.0$	0.1	$\pm(20 \% \text{ of reading})$
$30.1 \div 39.9$	0.1	$\pm(30 \% \text{ of reading})$

Test voltage frequency	125 Hz
Noise current indication	yes
Low clamp current indication	yes

Specific earth resistance

Measuring range (Ωm)	Resolution (Ωm)	Accuracy
$0.0 \div 99.9$	0.1	Calculated value, consider earth resistance 4-wire method
$100 \div 999$	1	
$1.00k \div 9.99k$	0.01k	
$10.0k \div 99.9k$	0.1k	
$>100k$	1k	

Wenner method principle with equal distances between test probes:
 $\rho = 2 \cdot \pi \cdot \text{distance} \cdot R$, with R as measured resistance in 4-wire method.

TRMS Clamp current

Measuring range (I)	Resolution (mA)	Accuracy
$0.0 \text{ mA} \div 99.9 \text{ mA}$	0.1mA	$\pm(3 \% \text{ of reading} + 3 \text{ digits})$
$100 \text{ mA} \div 999 \text{ mA}$	1 mA	$\pm(3 \% \text{ of reading} + 3 \text{ digits})$
$1.00 \text{ A} \div 19.99 \text{ A}$	0.01 A	$\pm(3 \% \text{ of reading} + 3 \text{ digits})$

Input resistance	100 Ω
Maximum input current	30 mA (=30 A @ current clamp with ratio 1000:1)
Measurement principle	current clamp, ratio 1000:1
Nominal frequency	40 Hz \div 500 Hz

General data

Power supply voltage	6 x 1.5 V battery or rechargeable battery, size AA
Operation	typical 20 h
Overvoltage category	50 V CAT IV
Protection classification	double insulation
Pollution degree	2
Protection degree	IP 40
Display	128 x 64 dots matrix display with backlight
Communication transfer speed	115200 baud (RS 232) / 256000 baud (USB)
Memory	1900 measurements
Dimensions (w x h x d)	14 cm x 8 cm x 23 cm
Weight without battery cells	0,85 kg

Ordering information

MI 3123

Standard set



Instrument MI 3123 Earth / Clamp

Soft hand strap

Earth test lead 4,5 m (blue)

Earth test lead 4,5 m (red)

Earth test lead 20 m (green)

Earth test lead 20 m (black)

Earth test rod, 4 pcs

Power supply adapter + 6 NiMH rechargeable AA batteries

Instruction manual on CD

Short instruction manual

Product verification data

Handbook "Guide for testing and verification of low voltage installations" on CD

Optional Accessories

Photo	Order No.	Description
	A 1018	Current clamp (low range, leakage)
	A 1019	Current clamp
	A 1160	Fast battery charger with a set of 6 pcs NiMH rechargeable AA batteries
	A 1271	Small soft carrying bag
	A 1289	Soft carrying bag
	A 1290	PC SW EuroLink PRO Plus with USB and RS232/PS2 cable
	A 1291	PC SW EuroLink PRO with USB and RS232/PS2 cable
	S 2007	Earth test set 50 m
	S 2009	Test lead set 2 m, 4 colours, 4 pcs



Measuring and Regulation Equipment Manufacturer

METREL d.d.

Ljubljanska 77

SI-1354 Horjul

Tel: + 386 (0)1 75 58 200

Fax: + 386 (0)1 75 49 226

E-mail: metrel@metrel.si

http://www.metrel.si

Note! Photographs in this catalogue may slightly differ from the instruments at the time of delivery.
Subject to technical change without notice.